

BrightSpeed 16, 8, 4

SCHEDULE A
GANTRY REVOLUTIONS:
DATE:
SYSTEM ID:

SITE:

<u>Procedure</u>	<u>Inspect</u>	<u>Clean</u>	<u>Split</u>
CONSOLE			
Console Inspection	_____	_____	_____
GANTRY			
Removal of Slip Ring Brush Debris	_____	_____	_____
Schleifring Maintenance and brush inspection	_____	_____	_____
PDU			
CPDU PM Checks	_____	_____	_____
NGPDU PM Checks	_____	_____	_____
System			
Verify Emergency Off Buttons	_____	_____	_____
Check Scan Control Pushbuttons	_____	_____	_____
Update Site Log	_____	_____	_____
Check Web for Latest PM	_____	_____	_____
Table			
Tighten Motor Coupling Screws	_____	_____	_____

Comments:

Is Follow Up Work Required ? Yes No

If yes, indicate the reschedule date: _____

Field Engineer:

Id#:

Dispatch#:

GE Medical Systems

PM REPORT

BrightSpeed 16, 8, 4

SCHEDULE A
GANTRY REVOLUTIONS:
DATE:
SYSTEM ID:

SITE:

Procedure

Inspect

Clean

Split

This Page Intentionally Left Blank.
