# *14HW30.x Load From Cold (LFC) Quick Guide*

| | | Approximate Time |
|---|---|---|
| **Preparations** | ❑ Disconnect Console from Hospital Network<br>❑ Remove all media from Peripheral Towers (DVD and MOD drives) | 2 min |
| **Save System State** | ❑ Perform Save System State, DVD-RAM: CSD > Utilities > System State<br>USB Media: CSD > Utilities > System State - USB<br>❑ Remove System State disk or USB Media and set aside | 10 min |
| **OS Load** | ❖ Place OS disk in Host DVD drive<br>❖ Shutdown and **Power-cycle** console (Power off, wait 30 secs, power on)<br>❖ Select "GEHC-CT" in boot process GUI.<br>❑ When system displays OS Load Complete window, remove OS disk and put Apps disk, then press [OK]. | 20 min |
| **Apps Load** | ❖ Click **Load** on CT Software Installation window<br>❖ Click **Yes** on Install INFO file window (Select ***No***, if missing System State Backup)<br>*Note: Manually configure system if missing System State Backup and Click **Accept** when finished*<br>❖ Insert System State DVD or USB Media in drive and click **OK**<br>❖ Click **Accept**, then ***Yes***, in INFO window, then click **OK**<br>❖ Shell window pops-up installing Apps software<br>❑ When finished, Host will automatically reboot, Close "Auto-Start Disabled" pop-up by clicking **OK**<br>❑ Remove Apps disk after reboot from Host DVD drive | 20 min |
| **LFC Load Menu**<br><br>NOTE:<br>Menu items will differ depending on System configuration | ❖ Open Terminal Window and log on as Root<br>❑ Type: **/usr/g/scripts/LFC/startLfcMenu.sh** LFC Menu appears<br><br>❖ Select **Set Date and Time** (1) and press **ENTER**<br>❖ Follow on screen instructions<br>❑ Close shell window when complete to return to LFC Menu<br><br>❖ Select **Install ClassC** (2). (Skip if Service Class C is not required)<br>❖ Insert ClassC CD in Host PC and press **ENTER** when complete<br>❑ Close shell window when complete to return to LFC Menu<br><br>❖ Select **Install ClassM** (3). (Skip if Service Class M is not required)<br>❖ Insert ClassM CD in Host PC and press **ENTER** when complete<br>❑ Close shell window when complete to return to LFC Menu<br><br>❖ Select **Restore System State** (4)<br>❖ Insert System State (System State DVD or USB Media saved earlier), press **ENTER**<br>*Note: If Options have been previous loaded and saved to System State, Options will be restored*<br>❑ Close shell window when complete to return to LFC Menu<br><br>❖ Select **Install Options** (5) and press **ENTER** (Skip if Options already restored)<br>*Note: Perform if missing System State Backup*<br>❑ Close shell window when complete to return to LFC Menu<br>❖<br>❑ Select **Quit** and press **ENTER**<br>❑ Confirm Quit by selection **Yes**<br>*Note: LFC Load Menu will no longer appear if Yes is selected* | 15 min<br>To<br>30 min |
| **TIC** | Following reboot, perform TIC if prompted<br>❑ Perform TIC procedure. CSD > Config. Tab > Tube Install Certification. | 2 min |
| **Flash Download** | Skip step if same revision software is being loaded on system<br>Only needed if first time or new software revision installs<br>❑ Perform Flash Download. CSD > Utilities > Flash Download<br>❑ Reconnect Console to Hospital Network<br>❖ **Reboot** | 20 min |
| **Functional Test** | ❑ Install Service Pack<br>❑ Perform "Save" System State. CSD > Utilities > System State<br>❑ Perform System Scanning Tests Procedure | 35 min |

| Last Updated | System | OS | APPS |
|---|---|---|---|
| September 12, 2014 | Optima CT660 | GEHC-SLES-11-SP1-Build6 | 14HW30.x |